



CERTIFICATE OF ANALYSIS

Nova Na/pH Conditioning Solution
Lot Number 26175005

Ref Number 00018
Exp Date 2027-09-24
Date of Manufacture 2026-06-24
UDI# (01)00385480000185(11)260624(17)270924(10)26175005

	<u>Spec</u>	<u>Pass/Fail</u>
pH	≤ 5.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Auishi-M

Approval

QC Inspector
Title

7/7/2026
Date