



CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 pH/Gas Reference Sensor
Lot Number 26183014

Part Number 57467
Exp Date 2028-01-02
Date of Manufacture 2026-07-02

	<u>Spec*</u>	<u>Pass/Fail</u>
pH Slope	8.70 - 11.80	Pass
BioPro Ctl L 1	6.605 - 6.705	Pass
BioPro Ctl L 3	7.319 - 7.419	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to test specifications.

* Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in purple ink, appearing to read "M. Houghton", with a long, sweeping horizontal line extending to the right.

Approval

QC Inspector
Title

07/08/2026
Date