



CERTIFICATE OF ANALYSIS

Stat Profile pHox Plus M Control

Ref Number 45205

Lot Number 26174019

Exp Date 2027-12-15

Date of Manufacture 2026-06-15

UDI# (01)00385480452052(11)260615(17)271215(10)26174019

Level 1 L/N: 26166093
Level 1 Exp Date: 2027-12-15
Spec Pass/Fail
pH 7.110-7.190 Pass
pCO2 mmHg 55.0-65.0 Pass
pO2 mmHg 58.0-68.0 Pass
Na+ mmol/L 157.8-161.8 Pass
K+ mmol/L 5.71-6.01 Pass
Ca++ mmol/L 1.46-1.54 Pass
Mg++ mmol/L 0.91-1.05 Pass
Glu mg/dL 73-87 Pass
Lac mmol/L 0.8-1.2 Pass

Level 2 L/N: 26166094
Level 2 Exp Date: 2027-12-15
Spec Pass/Fail
pH 7.350-7.420 Pass
pCO2 mmHg 39.0-45.0 Pass
pO2 mmHg 97.0-107.0 Pass
Na+ mmol/L 140.0-144.0 Pass
K+ mmol/L 3.76-3.96 Pass
Ca++ mmol/L 0.94-1.02 Pass
Mg++ mmol/L 0.52-0.60 Pass
Glu mg/dL 185-215 Pass
Lac mmol/L 2.5-3.1 Pass

Level 3 L/N: 26166095
Level 3 Exp Date: 2027-12-15
Spec Pass/Fail
pH 7.570-7.650 Pass
pCO2 mmHg 18.0-26.0 Pass
pO2 mmHg 135.0-149.0 Pass
Na+ mmol/L 118.0-122.0 Pass
K+ mmol/L 1.85-2.15 Pass
Ca++ mmol/L 0.50-0.60 Pass
Mg++ mmol/L 0.28-0.36 Pass
Glu mg/dL 285-325 Pass
Lac mmol/L 6.4-7.7 Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

Qc Inspector
Title

7/7/2026
Date