



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 1

Nova Ref Number 41741

LifeScan Ref Number 02307701

Lot Number 0426133301

Exp Date 2028-11-13

Date of Manufacture 2026-05-13

UDI # (01)00385480417419(11)260513(17)281113(10)0426133301

	Spec	Mean Result
Glucose mg/dL	57 - 65	61

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi.M

Approval

QC Inspector

Title

6/30/2026

Date