



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 200 Sample
Lot Number 26166091

Ref Number 53360
Exp Date 2027-11-29
Date of Manufacture 2026-05-29
UDI# (01)00385480533607(11)260529(17)271129(10)26166091

Cal A	L/N	26149024		
			<u>Assigned Value*</u>	<u>Result</u>
pH			7.354	N/A
pCO2 mmHg			25.0	N/A

Cal B	L/N	26149025		
			<u>Spec</u>	<u>Pass/Fail</u>
pH			6.837 - 6.843	Pass
pO2 mmHg			167.0	N/A

Flush	L/N	26149027		
			<u>Assigned Value*</u>	<u>Result</u>
pCO2 mmHg			52.2	N/A
pO2 mmHg			96.2	N/A

Reference	L/N	26149026		
			<u>Spec</u>	<u>Pass/Fail</u>
KCI			1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi.M

Approval

QC Inspector

Title

07/02/2026

Date