



CERTIFICATE OF ANALYSIS

STP pHox Ultra / CCX Calcium Sensor
Lot Number 26175023

Part Number 48581
Exp Date 2027-09-24
Date of Manufacture 2026-06-24
UDI# (01)00385480485814(11)260624(17)270924(10)26175023

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	8.50 - 12.50	Pass
Std C mV	0.00 $\pm$ 50.00	Pass
pHox Ctl L4	1.03 - 1.19	Pass
pHox Ctl L5	1.51 - 1.75	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Khachaturian", written over a faint, light blue grid background.

Approval

QC Inspector
Title

7/14/2026
Date