



CERTIFICATE OF ANALYSIS

Reagent Pack Autosampler Flex 2
Lot Number 26163043

Part Number 62450
Exp Date 2027-06-02
Date of Manufacturing 2026-06-02

Solution DP L/N	26153046
<u>Spec</u>	<u>Pass/Fail</u>
NaOCl % 0.43 - 0.73	Pass

Solution SF L/N	26155035
<u>Spec</u>	<u>Pass/Fail</u>
pH 6.900 - 7.100	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: Na, HCO₃, pH, KCl, iCa, NH₄, K.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi-M

Approval

QC Inspector
Title

7/15/2026
Date