



CERTIFICATE OF ANALYSIS

BioProfile pHox Auto-Cartridge QC
Lot Number 26166097

Ref Number 42751
Exp Date 2027-12

Level 1			Level 2		
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
pH	6.600 - 6.800	Pass	pH	6.750 - 6.950	Pass
PCO2 mmHg	57.0 - 67.0	Pass	PCO2 mmHg	37.0 - 43.0	Pass
PO2 mmHg	53.0 - 63.0	Pass	PO2 mmHg	93.0 - 103.0	Pass

Level 3		
	<u>Spec</u>	<u>Pass/Fail</u>
pH	7.200 - 7.400	Pass
PCO2 mmHg	17.5 - 22.5	Pass
PO2 mmHg	133.0 - 143.0	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. The following analytes are traceable

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi.M

Approval

QC Inspector

Title

7/17/2026

Date