



CERTIFICATE OF ANALYSIS

STP Prime Plus Reference Cartridge
Lot Number 26175055

Ref Number 57823
Exp Date 2027-12-24
Date of Manufacture 2026-06-24
UDI# (01)00385480578233(11)260624(17)271224(10)26175055

Internal QC Level 4	<u>Spec</u>	<u>Pass/Fail</u>
Na, mmol/L	136.7 - 144.7	Pass
K, mmol/L	3.69 - 4.19	Pass
CL, mmol/L	122.0 - 131.0	Pass
External QC Level 4	<u>Spec</u>	<u>Pass/Fail</u>
Na, mmol/L	134.3 - 142.3	Pass
K, mmol/L	3.73 - 4.23	Pass
CL, mmol/L	118.9 - 127.9	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2016, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in purple ink, appearing to read "M. Houghton", with a long, sweeping horizontal line extending to the right.

Approval

QC Inspector
Title

7/14/2026
Date