



CERTIFICATE OF ANALYSIS

STP pHox Plus/L/C Chloride Sensor
Lot Number 26175022

Part Number 27812

Exp Date 2027-09-24

Date of Manufacture 2026-06-24

UDI# (01)10385480278123(11)260624(17)270924(10)26175022

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	≥ 9.00	Pass
Std A mV	55.00 - 95.00	Pass
pHOx Ctl L1	119.9 - 128.9	Pass
pHOx Ctl L3	78.2 - 88.2	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Khankhyan", written over a faint, light blue rectangular stamp.

Approval

QC Inspector

Title

7/16/2026

Date