



CERTIFICATE OF ANALYSIS

Nova StatStrip Lactate Control Solution Level 1
Lot Number 2726162301

Ref Number 47553

Exp Date 2028-12-11

Date of Manufacture 2026-06-11

UDI# (01)00385480475532(11)260611(17)281211(10)2726162301

	<u>Spec</u>	<u>Mean Result</u>
Lactate mmol/L	0.5-0.7	0.6

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "J. M. ...", written over a horizontal line.

Approval

QC Inspector
Title

7/16/2026
Date