



## CERTIFICATE OF ANALYSIS

BioProfile FLEX 2 pH/Gas Reference Sensor  
Lot Number 26188042

Part Number 57467  
Exp Date 2028-01-07  
Date of Manufacture 2026-07-07

|                | <u>Spec*</u>  | <u>Pass/Fail</u> |
|----------------|---------------|------------------|
| pH Slope       | 8.70 - 11.80  | Pass             |
| BioPro Ctl L 1 | 6.638 - 6.738 | Pass             |
| BioPro Ctl L 3 | 7.319 - 7.419 | Pass             |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to test specifications.

\* Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in purple ink, appearing to read "M. Houghton", with a long, sweeping horizontal line extending to the right.

Approval

QC Inspector  
Title

07/17/2026  
Date