



CERTIFICATE OF ANALYSIS

Prime Sensor Card ABG
Lot Number 26202045

Catalog Number 42031
Exp Date 2027-03-14
Date of Manufacture 2026-07-14
UDI# (01)10385480420317(11)260714(17)270314(10)26202045

	<u>Slope</u>	<u>Pass/Fail</u>
pH	8.50 - 12.00	Pass
PCO2	1.00 - 8.00	Pass
PO2	2.80 - 12.20	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Khankh", written over a faint, rectangular grid background.

Approval

QC Inspector
Title

7/23/2026
Date