



CERTIFICATE OF ANALYSIS

StatStrip Glucose Control Level 3

Nova Reference 41743
LifeScan Reference 02307901
Lot Number 0426183303

Exp Date 2029-01-02
Date of Manufacture 2026-07-02
UDI#(01)00385480417433(11)260702(17)290102(10)0426183303

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275 - 325	299

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to read "M. Houghton", written over a horizontal line.

Approval

QC Inspector
Title

7/23/2026
Date