



CERTIFICATE OF ANALYSIS

STP pHox Ultra / CCX Chloride Sensor
Lot Number 26189017

Part Number 48557
Exp Date 2027-10-08
Date of Manufacture 2026-07-08
UDI# (01)10385480485576(11)260708(17)271008(10)26189017

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	8.50 - 13.00	Pass
pHOx Ctl L4	121.3 - 130.3	Pass
pHOx Ctl L5	91.5 - 100.5	Pass
Std C mV	10.00 - 90.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "C. S. V. A.", written over a horizontal line.

Approval

QC Inspector
Title

7/22/2026
Date