



## CERTIFICATE OF ANALYSIS

STP pHox Ultra / CCX Ionized Mg Sensor  
Lot Number 26182043

Part Number 48556  
Exp Date 2027-10-01  
Date of Manufacture 2026-07-01  
UDI# (01)10385480485569(11)260701(17)271001(10)26182043

|             | <u>Spec</u>      | <u>Pass/Fail</u> |
|-------------|------------------|------------------|
| Slope       | 9.70 - 14.00     | Pass             |
| Std C mV    | -100.00 - +60.00 | Pass             |
| pHox Ctl L4 | 0.55 - 0.73      | Pass             |
| pHox Ctl L5 | 1.00 - 1.30      | Pass             |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "C. S. S. S.", written over a horizontal line.

Approval

QC Inspector  
Title

7/21/2026  
Date