



CERTIFICATE OF ANALYSIS

Nova Lactate Plus Control Solution Level 1
Lot Number 3826100301

Ref Number 40815

Exp Date 2028-10-10

Date of Manufacture 2026-04-10

UDI# (01)00385480408158(11)260410(17)281010(10)3826100301

	<u>Spec</u>	<u>Mean Result</u>
Lactate mmol/L	1.3 - 1.5	1.3

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "J. Smith", written over a horizontal line.

Approval

QC Inspector

Title

6/12/2026

Date