

OI-922

Page 28 of 67

Rev. BS

Vialed and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	STATSENSOR-I CREAT Test Strips
Catalog Number:	P/N 43272
Nova Biomedical P/N	P/N 43272
Control Level 1: CV (%) $\leq$ 15% Control Level 2: CV (%) $\leq$ 10% Control Level 3: CV (%) $\leq$ 10%	PASS / FAIL PASS / FAIL PASS / FAIL
Qty:	6,156
Lot No.	4926194129
Mfg. Date:	2026-07-13
Expiration Date:	2027-07-13
Unique Device Identifier (UDI):	(01)20385480432720 (11) <u>26</u> <u>07</u> <u>13</u> (17) <u>27</u> <u>07</u> <u>13</u> (10) <u>4926194129</u>

Maureen Broome
Quality Control Manager

7/27/26
Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454