



CERTIFICATE OF ANALYSIS

Nova StatStrip Hb/Hct Control Level 2
Lot Number 2026177242

Ref Number 59827
Exp Date 2028-06-26
Date of Manufacture 2026-06-26
UDI# (01)00385480598279(11)260626(17)280626(10)2026177242

	<u>Spec</u>	<u>Mean Result</u>
Hb g/dL	14.2 - 15.8	14.7
Hct %	42 - 47	42

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi.M

Approval

QC Inspector
Title

7/23/2026
Date