



CERTIFICATE OF ANALYSIS

Nova StatStrip Hb/Hct Control Level 1
Lot Number 2026162241

Ref Number 59820
Exp Date 2028-06-11
Date of Manufacture 2026-06-11
UDI#(01)00385480598200(11)260611(17)280611(10)2026162241

	<u>Spec</u>	<u>Mean Result</u>
Hb g/dL	8.2 - 8.8	8.6
Hct %	24 - 26	26

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi.M

Approval

QC Inspector
Title

7/24/2026
Date