

OI-922

Page 32 of 67

Rev. BS

Vial and Packaged Test Strip Release

Test Strips

nova
biomedical

CERTIFICATE OF COMPLIANCE

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	NIA
Description:	TEST STRIPS, NOVA MAX PRO CREATININE eGFR, 50 COUNT
Catalog Number:	63983
Nova Biomedical P/N	63983
Control Level 1: CV (%) $\leq$ 15% Control Level 2: CV (%) $\leq$ 10%	PASS / FAIL PASS / FAIL
Qty:	572
Lot No.	2730226198
Mfg. Date:	2026-07-17
Expiration Date:	2027-07-17
Unique Device Identifier (UDI):	(01)20385480639839 (11)260717 (17)270717 (10)2730226198

Marc Browne

Quality Control Manager

7/29/26
Date

Nova Biomedical Corporation

200 Prospect Street Waltham, MA 02454