



CERTIFICATE OF ANALYSIS

BioProfile Flex2 Performance Check ESM
Lot Number 26201004

Ref Number 60095
Exp Date 2027-07-20
Date of Manufacture 2026-07-20

	<u>Spec</u>	<u>Pass/Fail</u>
pH	7.400 - 7.450	Pass
Gluc g/L	4.50 - 5.50	Pass
Osmo mOsm/kg H2O	270 - 290	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with ISO 9001:2008, Quality Management System Requirements and current Good Manufacturing Practices (cGMP), and conforms to the indicated test specifications. The following analytes are traceable to NIST SRM Materials: pH

A handwritten signature in black ink, appearing to be "J. Smith", written over a horizontal line.

Approval

QC Inspector
Title

7/28/2026
Date