



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 2
LifeScan Ref Number 02307801

Ref Number 41742

Lot Number 0426194302

Exp Date 2029-01-13

Date of Manufacture 2026-07-13

UDI# (01)00385480417426(11)260713(17)290113(10)0426194302

	Spec	Mean Result
Glucose mg/dL	105-125	113

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in purple ink, appearing to read "M. Houghton".

Approval

QC Inspector

Title

7/29/2026

Date