



CERTIFICATE OF ANALYSIS

Preheater Deproteinizing Sol'n
Lot Number 26182006

Ref Number 12704
Exp Date 2027-07-01
Date of Manufacture 2026-07-01
UDI# (01)00385480127042(11)260701(17)270701(10)26182006

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl	0.70 - 0.90%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in black ink, appearing to be "J. M. ...", written over a horizontal line.

Approval

QC Inspector
Title

7/28/2026
Date