



CERTIFICATE OF ANALYSIS

Allegro Lipids Test Cartridge
Lot Number 26201014

Ref Number 54652
Exp Date 2028-01-20
Date of Manufacture 2026-07-20
UDI# (01)00385480546522(11)260720(17)280120(10)26201014

Ctl Level 1				Ctl Level 2			
	<u>Spec</u>		<u>Pass/Fail</u>		<u>Spec</u>		<u>Pass/Fail</u>
HDL mg/dL	26	- 48	Pass	HDL mg/dL	71	- 105	Pass
TCHOL mg/dL	146	- 216	Pass	TCHOL mg/dL	286	- 402	Pass
TGL mg/dL	76	- 152	Pass	TGL mg/dL	310	- 450	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

7/30/2026
Date