



CERTIFICATE OF ANALYSIS

STP pHox Plus/L/C/M Potassium Sensor
Lot Number 26196041

Part Number 27811

Exp Date 2027-10-15

Date of Manufacture 2026-07-15

UDI# (01)00385480278119(11)260715(17)271015(10)26196041

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	≥ 9.60	Pass
Std A mV	0 ± 60.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Sean Quinn".

Approval

QC Inspector

Title

07/30/2026

Date