



CERTIFICATE OF ANALYSIS

Stat Profile pHox Ultra / CCX BUN Membrane Kit
Lot Number 26204002

Part Number 48596
Exp Date 2027-10-23

Date of manufacture 2026-07-23

UDI# (01)00385480485968(11)260723(17)271023(10)26204002

	<u>Spec</u>	<u>Mean Result</u>
BUN Slope	9.30 - 12.90	10.64
CCX L 4	14 - 24	39
CCX L 5	49 - 63	39

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "J. M. ...", written over a faint horizontal line.

Approval

QC Inspector
Title

8/3/2026
Date