



## CERTIFICATE OF ANALYSIS

STP Prime Plus Reference Cartridge  
Lot Number 26205086

Ref Number 57823  
Exp Date 2028-01-24  
Date of Manufacture 2026-07-24  
UDI# (01)00385480578233(11)260724(17)280124(10)26205086

| Internal QC Level 4 | <u>Spec</u>   | <u>Pass/Fail</u> |
|---------------------|---------------|------------------|
| Na, mmol/L          | 136.7 - 144.7 | Pass             |
| K, mmol/L           | 3.69 - 4.19   | Pass             |
| CL, mmol/L          | 122.0 - 131.0 | Pass             |
| External QC Level 4 | <u>Spec</u>   | <u>Pass/Fail</u> |
| Na, mmol/L          | 134.3 - 142.3 | Pass             |
| K, mmol/L           | 3.73 - 4.23   | Pass             |
| CL, mmol/L          | 118.9 - 127.9 | Pass             |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2016, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Khankhyan", written over a faint, rectangular grid background.

Approval

QC Inspector  
Title

8/3/2026  
Date