



CERTIFICATE OF ANALYSIS

STP pHox Plus M ionized Magnesium Sensor
Lot Number 26196038

Part Number 45172
Exp Date 2027-10-15
Date of Manufacture 2026-07-15
UDI# (01)10385480451724(11)260715(17)271015(10)26196038

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.70 - 14.00	Pass
Std A mV	-100.00 - +60.00	Pass
pHOx + M CtL 1	0.83 - 1.13	Pass
pHOx + M CtL 3	0.26 - 0.38	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Sean Quinn".

Approval

QC Inspector
Title

07/29/2026
Date