



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Glucose Membrane
Lot Number 26194035

Part Number 48552
Exp Date 2027-10-13

Date of Manufacture 2026-07-13

UDI# (01)00385480485524(11)260713(17)271013(10)26194035

	<u>Spec*</u>	<u>Mean Result</u>
Slope	11.00 - 45.00	37.22
pHox Ultra/CCX Ctl L4	73 - 87	80
pHox Ultra/CCX Ctl L5	255 - 305	290

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Sean Quinn".

Approval

QC Inspector
Title

8/5/2026
Date