



CERTIFICATE OF ANALYSIS

NCRT 10 Calibrator Pack
Lot Number 619410

Ref Number 14608
Exp Date 2028-01-13
Date of Manufacture 2026-07-13
UDI# (01)00385480146081(11)260713(17)280113(10)619410

Cal A	L/N	26194010		Cal B	L/N	26194010	
		<u>Spec</u>	<u>Pass/Fail</u>			<u>Spec</u>	<u>Pass/Fail</u>
Na	mmol/L	139.0 - 141.0	Pass	Na	mmol/L	59.4 - 60.6	Pass
K	mmol/L	3.97 - 4.03	Pass	K	mmol/L	9.8 - 10.2	Pass
Cl	mmol/L	113.6 - 115.4	Pass	Cl	mmol/L	45.5 - 46.5	Pass
TCa	mmol/L	1.94 - 2.06	Pass	TCa	mmol/L	3.88 - 4.12	Pass
Li	mmol/L	0.50 - 0.51	Pass	Li	mmol/L	4.95 - 5.05	Pass

Ref L/N 26194010

	<u>Spec</u>	<u>Result</u>
Ca++ Releasing Agent	N/A	Present

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Awishi.M

Approval

QC Inspector
Title

8/24/2026
Date