



CERTIFICATE OF ANALYSIS

Nova 8 Ionized Magnesium Sensor
Lot Number 26219007

Ref Number 15265

Exp Date 2027-11-07

Date of Manufacture 2026-08-07

UDI# (01)10385480152652(11)260807(17)271107(10)26219007

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	10.5 - 13.5	Pass
Chem Control L1 mmol/L	0.39 - 0.59	Pass
Set A Level 2 mmol/L	0.78 - 1.02	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Sean Quinn".

Approval

QC Inspector

Title

09/01/2026

Date