



CERTIFICATE OF ANALYSIS

NCRT 4 Calibrator Pack
Lot Number 623915

Ref Number 15647
Exp Date 2027-11-14
Date of Manufacture 2026-08-14
UDI# (01)00385480156479(11)260814(17)271114(10)623915

CAI A	L/N	26238002	Cal B	L/N	26238003	Cal H	L/N	26226048	
		<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Result</u>
Na mmol/L	129.1 - 130.9	Pass	Na mmol/L	59.4 - 60.6	Pass	CO2 Releasing Agent	N/A	Present	
K mmol/L	3.97 - 4.03	Pass	K mmol/L	9.8 - 10.2	Pass				
Cl mmol/L	115.1 - 116.9	Pass	Cl mmol/L	60.4 - 61.6	Pass				
HCO3 mmol/L	19.2 - 20.8	Pass	HCO3 mmol/L	9.6 - 10.4	Pass				
Li mmol/L	0.49 - 0.51	Pass	Li mmol/L	4.95 - 5.05	Pass				
			Reference	L/N	26226047				
					<u>Spec</u>	<u>Pass/Fail</u>			
			KCl mol/L	1.8 - 2.2	Pass				

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02453 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Auishi.M

Approval

QC Inspector
Title

9/9/2026
Date