



CERTIFICATE OF ANALYSIS

Stat Profile® pHox Ultra™/CCX™ Ionized Mg Sensor
Lot Number 23030055

Ref Number 48556

Exp Date 2024-04-30

Date of Manufacture 2023-01-30

UDI# (01)10385480485569(11)230130(17)240430(10)23030055

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.70 to 14.00	Pass
Std C mV	-100.00 - +60.00	Pass
CCX L4	0.55 - 0.73	Pass
CCX L5	1.00 - 1.30	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

3/27/2023

Date