



CERTIFICATE OF ANALYSIS

Nova Control Level 3
Lot Number 23030032

Ref Number 18654
Level 3 Exp Date 2025-01-30
Date of Manufacture 2023-01-30

	<u>Spec</u>	<u>Pass/Fail</u>
Na+ mmol/L	145-171	Pass
K+ mmol/L	1.7-2.3	Pass
Cl- mmol/L	105-127	Pass
TCO2 mmol/L	28-40	Pass
Glu mg/dL	312-408	Pass
BUN mg/dL	21-35	Pass
TCa mg/dL	2.68-4.31	Pass
iCa++ mg/dL	2.08-3.31	Pass
iMg++ mg/dL	0.74-1.23	Pass
Li+ mmol/L	2.59-3.50	Pass
pH	7.25-7.50	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector
Title

3/24/2023
Date