



CERTIFICATE OF ANALYSIS

STAT PROFILE® pHox Ultra™/CCX™ pO2 Sensor
Lot Number 23055028

Ref Number 48598
Exp Date 2025-02-24
Date of Manufacture 2023-02-24
UDI# (01)00385480485982(11)230224(17)250224(10)23055028

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	-3.00 to -6.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

4/7/2023
Date