

OI-922

Page 28 of 66

Rev. BH

Vialed and Packaged Test Strip Release

Test Strips

*nova*  
biomedical

**CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

Customer:	N/A
Description:	STATSENSOR-I CREAT Test Strips
Catalog Number:	P/N 43272
Nova Biomedical P/N	P/N 43272
Control Level 1: CV (%) $\leq$ 15% Control Level 2: CV (%) $\leq$ 10% Control Level 3: CV (%) $\leq$ 10%	PASS / FAIL PASS / FAIL PASS / FAIL
Qty:	4,266
Lot No.	4923108129
Mfg. Date:	2023-04-18
Expiration Date:	2024-04-18
Unique Device Identifier (UDI):	(01)20385480432720 (11)230418 (17)240418 (10)4923108129

*Marc Broome*  
Quality Control Manager

*4/27/2023*  
Date

Nova Biomedical Corporation

200 Prospect Street

Waltham, MA 02454