



CERTIFICATE OF ANALYSIS

STAT PROFILE pHox Plus Glucose Sensor
Lot Number 22160047

Part Number 27809
Exp Date 2024-06-09

Date of Manufacture 2022-06-09

UDI# (01)00385480278096(11)220609(17)240609(10)22160047

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	11.00 - 45.00	Pass
pHOx Ctl L3	75 - 89	Pass
Set A L1	≥ 470	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

7/1/2022
Date