



## CERTIFICATE OF ANALYSIS

STAT PROFILE pHox Plus Glucose Sensor  
Lot Number 23004028

Part Number 27809  
Exp Date 2025-01-04  
Date of Manufacture 2023-01-04  
UDI# (01)00385480278096(11)230104(17)250104(10)23004028

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	11.00 - 45.00	Pass
pHOx Ctl L3	76 - 90	Pass
Set A L1	$\geq 470$	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

2/9/2023  
Date