



CERTIFICATE OF ANALYSIS
StatStrip Glucose Control Solution Level 3

Nova Ref Number 54522
LifeScan Ref Number 02315602
Lot Number 0423110243

Date of Manufacture 2023-04-20
Exp Date 2025-04-20

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	275 - 325	297

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

05/05/2023
Date