



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 2

Nova Reference 54521
LifeScan Reference 02315502
Lot Number 0423110242
Date of Manufacture 2023-04-20
Exp Date 2025-04-20
UDI# (01)00385480545211(11)230420(17)250420(10)0423110242

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	105 - 125	116

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector

Title

05/05/2023

Date